OAJAE



Vol. 15, No. 4, 2021139, December 2021 doi: https://doi.org/10.5572/ajae.2021.139 ISSN (Online) 2287-1160, ISSN (Print) 1976-6912

Research Article

The Comparisons of Real-time Ammonia Adsorption Measurement in Varying Inlet Tubes and the Different Ammonia Measurement Methods in the Atmosphere

Kyunghoon Kim, Gyutae Park¹⁾, Seokwon Kang, Rahul Singh, Jeongin Song, Siyoung Choi, Inseon Park, Dong-Gil Yu, Myeong-Bok Kim, Min-Suk Bae²⁾, Suna Jung¹⁾, YuWoon Chang¹⁾, Jonghun Park³⁾, Hae-Jin Jung¹⁾, Yong-jae Lim^{1),*} and Taehyoung Lee^{*}

Department of Environmental Science, Hankuk University of Foreign Studies, Yongin, Republic of Korea ¹⁾Department of Climate and Air Quality, National Institute of Environmental Research, Incheon, Republic of Korea ²⁾Department of Environmental Engineering, Mokpo National University, Muan, Republic of Korea ³⁾APM Engineering Co., Ltd., Bucheon, Republic of Korea

*Corresponding author.

Tel: +82-32-560-7264 (Y. Lim) +82-31-330-4039 (T. Lee) E-mail: limyongjae@korea.kr (Y. Lim) thlee@hufs.ac.kr (T. Lee)

Received: 26 November 2021 Revised: 8 December 2021 Accepted: 15 December 2021 **ABSTRACT** Ammonia (NH₃) is an important, albeit sticky, precursor for producing secondary inorganic aerosols (SIA), especially in the form of ammonium nitrate (NH_4NO_3) and ammonium sulfate ($(NH_{d})_{2}SO_{d}$). To reduce SIAs, many researchers have attempted to measure the concentration of ambient NH₃ using real-time or passive methods. However, NH_3 is a highly sticky gas and is therefore difficult to measure using real-time methods without incurring losses during measurement. In this study, four different tubing materials, semi seamless tubes, perfluoroalkoxy (PFA), polytetrafluoroethylene (PTFE), and polyvinylidene fluoride (PVDF), were used to ascertain the adsorption of NH_3 in inlets using real-time instruments. Without heating sample tubes and at 0% relative humidity (RH), this study shows that PTFE had the least adsorption (i.e., 0% at 1 and 2 m of sample tube), and semi-seamless tubes had the highest adsorption (i.e., 27.5% at 1 m of sample tube). To calculate the adsorption of NH_3 under ambient conditions, at various inlet lengths, the RH of NH₃ was varied from 20% to 80%, which showed that shorter inlets and higher RH lower NH₃ adsorption at inlets (i.e., 1.74 ppb m⁻¹ at 80% RH and 7.48 ppb m⁻¹ at 20% RH). Additionally, inlet heating was effective in reducing the adsorption of NH₃ as the RH decreased. Applying the inlet system (i.e., 2 m of PTFE tube with heating) showed excellent correlation (slope: 0.995 and coefficient: 0.992) between two different real-time measurements while measuring ambient air.

KEY WORDS Ammonia, Inlet adsorption, Polytetrafluoroethylene, Protocol, Ammonia measurement

1. INTRODUCTION

Ammonia (NH₃) is a gaseous substance with very high adsorption properties (Bobrutzki *et al.*, 2010). It is emitted into the atmosphere via various activities such as excreta management, agriculture, and substances such as mobile pollutants (Paulot *et al.*, 2014). NH₃ emissions in South Korea have gradually increased from 292 Gg in 2014 to 301 Gg in 2016 (Park *et al.*, 2019). When NH₃ reacts with nitric acid (HNO₃), sulfuric acid (H₂SO₄), or hydrochloric acid (HCl),

This is an open-access article distributed under the terms of the Creative Commons Attribution Non-Commercial License (http://creativecommons. org/licenses/by-nc/4.0/), which permits unrestricted non-commercial use, distribution, and reproduction in any medium, provided the original work is properly cited. which are all gaseous substances, they transform into ammonium nitrate (NH₄NO₃), ammonium sulfate ((NH₄)₂SO₄), or ammonium chloride (NH₄Cl), respectively. These are important components for producing secondary inorganic aerosols (SIAs) (Park *et al.*, 2019; Hsieh and Chen, 2010; Sharma *et al.*, 2009). Such reactions result in reduced visibility and adverse effects on the human body (Kang *et al.*, 2020). Among the particulate matter currently found in South Korea, SIAs account for more than 50% of secondary particulate matter (Kim *et al.*, 2020; Park *et al.*, 2018), whereas NH₄NO₃ and (NH₄)₂SO₄ account for 35–62% of all secondary particulate matter (Jordan *et al.*, 2020).

Since NH₃ was identified as a major component of SIA formation, studies on the distribution of NH₃ in the atmosphere have been conducted worldwide. In the United States of America, NH₃ concentrations in various regions of the country have been measured since 2007 using passive samplers that capture NH₃ over a long time through the Ammonia Monitoring Network (AMoN) project of the National Atmospheric Deposition Program (NADP) (Zhou et al., 2015). China has also studied the concentration of NH₃ in its atmosphere using passive samplers (Pan et al., 2018). Although the passive sampler method has the advantage of being capable of simultaneously measuring NH₃ concentrations at various locations, it has the disadvantage of being unable to identify changes in concentration over time. Research on changes in NH₃ concentration over time, using real-time NH₃ measurement equipment, is being actively conducted (Zhou et al., 2019). This real-time measurement is disadvantageous as it makes it difficult to accurately measure NH₃ concentrations due to underestimation resulting from a loss in the inflow tube, attributed to a high NH₃ adsorption force (Yi et al., 2021; Bobrutzki et al., 2010). The real-time measurement of black carbon also has disadvantages at low concentration in some cases (Lee, 2019). Research is being actively conducted on these real-time measurements to resolve the underestimation (Pollak et al., 2019; Ellis et al., 2010). However, there is a lack of research on the degree of NH₃ adsorption onto the inflow tube and the resultant reduced NH₃ concentration when measuring atmospheric NH₃ levels.

Therefore, this study aimed to investigate the degree to which NH_3 is adsorbed onto inlet tubes using realtime NH_3 analysis equipment. Additionally, the degree to which NH_3 is adsorbed onto the tube was studied according to moisture, tube length, and heating status after tubes made of different materials were selected, with the ultimate objective of identifying a material that is most suitable for use in actual measurements. Furthermore, the "most suitable" experimental tube identified here is defined as an inlet tube for real-time analysis equipment, to compare the actual atmospheric NH_3 concentrations and the NH_3 concentrations measured at one-week intervals using NH_3 adsorption and collection methods.

2. EXPERIMENT

2.1 Measurement Site

In this study, a comparison of the properties of NH₃ adsorption according to moisture, inlet tube length, inlet tube material, and heating status was carried out at the Natural Science College on the Global Campus at the Hankuk University of Foreign Studies, Cheoin-gu, Yongin-si, Gyeonggi-do, South Korea (37.34°N, 127.27°E). To further identify the differences in NH₃ concentrations between real-time analysis equipment and the method of analyzing NH₃ after long-term data collection, atmospheric NH₃ concentrations were measured from May 19, 2020, to July 12, 2020, and from December 4, 2020, to January 3, 2021, at the Honam Air and Environment Research Institute located in Gwangju, South Korea (35.23°N, 126.85°E). The Honam Air and Environment Research Institute is located northwest of the Gwangju Metropolitan City, with agricultural land near the measuring station and residential complexes concentrated to the south side of the measuring station.

2.2 Method for Measuring NH₃ Adsorption Depending on Moisture and Inlet Sampling Conditions

Real-time NH₃ measurements, according to humidity and heating of inlet tubes, were analyzed at intervals of 1 s using the G2103 (Picarro, California, USA), based on the principle of cavity ring-down spectroscopy (CRDS). Standard NH₃ gas at a concentration of 1.001 ppm (Rigas, Daejeon, Korea) and purified zero air (Union Gas, Gyeonggi-do, Korea) were diluted to 100 ppb using a mass flow controller (MKP, Gyeonggi-do, Korea) to measure atmospheric NH₃ concentrations. Moreover, by installing an additional needle valve and a bubbler in the tube that connects to zero air, the moisture content in the sample can be adjusted, as zero air can contain mois-



Fig. 1. Schematic diagram of NH_3 adsorption test. The bubbler flow and composition of the test tube (red line) were changed after each measurement.

ture. All background tubes through which the gases were diluted and passed through were made of PTFE, with an outer diameter of 6.35 mm and valves divided from the background tube to the sample tube (made of Teflon) to minimize NH_3 adsorption. To measure the amount of moisture produced through the bubbler, HYGROLOG NT (Rotronic AG, Switzerland), a temperature and moisture meter, were additionally installed in a discharge tube without entering the equipment (see Fig. 1).

The various tube materials used to investigate the degree of NH₃ adsorption, according to the material of the inlet tube, were semi-seamless tubes (SUS316L) made out of perfluoroalkoxy (PFA), polytetrafluoroethylene (PTFE), and polyvinylidene fluoride (PVDF) (see Table 1) with an outer diameter of 6.35 mm to minimize the difference due to the tube diameter. Furthermore, a tube made of PTFE (outer diameter: 6.35 mm) was divided into 1, 2, 3, and 4 m lengths to study the degree of NH₃ adsorption according to tube length. The relative humidity (i.e., 20, 40, 60, and 80%) of the injected sample was altered to determine the effect of moisture on adsorption (see Table 2). In all experiments conducted in this study, a waiting time of 60 min was spent both before and after the experiment for each tube to stabilize the background tube at a specific injection concentration, temperature, and humidity. Moreover, to keep the experimental tubes as clean as possible, they were cleaned from inside using N_2 (Air Korea) for 2 h both before and after use.

Table 1	. Materials	used durir	ng NH ₃	adsorption	test
---------	-------------	------------	--------------------	------------	------

Tube material	Manufacturer	Outer diameter (mm)	Inner diameter (mm)
Semi seamless tube	Hohwon	6.35	4
Perfluoro alkoxy	iXAK	6.35	4
Polytetrafluoroethyelene	iXAK	6.35	4.4
Polyvinylidene fluoride	Masterflex	6.35	2.6
Semi seamless tube Perfluoro alkoxy Polytetrafluoroethyelene Polyvinylidene fluoride	Hohwon iXAK iXAK Masterflex	(mm) 6.35 6.35 6.35 6.35	(mm) 4 4.4 2.6

To investigate the adsorptive properties of NH_3 , the area of adsorption was reduced until NH_3 concentrations were reduced after changing NH_3 concentrations using the experimental tube, which then returned to its original concentration before the induced change. These data were then used to calculate the ratio of lost NH_3 , as shown in Fig. 2 (Vaittinen *et al.*, 2014).

2.3 Measurement of Atmospheric NH₃ Using Real-time and Non-real-time Measurement Methods

For the measurement of atmospheric NH_3 concentrations, real-time analysis was carried out using the EAA-30r-EP (Los Gatos Research, California, USA) from May 2020 to July 2020 and from December 2020 to January 2021; used alongside G2103 in December 2020, the samples from this equipment were collected using annular denuders (URG Corporation, USA) and passive 🔘 AJAE Asian Journal of Atmospheric Environment, Vol. 15, No. 4, 2021139, 2021

Test type	Tube material	Length (m)	Heating	RH(%)
Material	Semi seamless tube	1	No	0
	Perfluoro alkoxy	1, 2	No	0
	Polytetrafluoroethylene	1, 2	No	0
	Polyvinylidene fluoride	1, 2	No	0
Length and RH	Polytetrafluoroethylene	1, 2, 3, 4	No	20, 40, 60, 80
Heating	Polytetrafluoroethylene	2	Yes, No	20, 40, 60, 80





Fig. 2. Real-time measurement of NH_3 loss used to calculate adsorption ratio. The red area indicates the loss of NH_{3} , and black indicates the measured concentrations of NH_3 .

samplers (Radiello, Pavia, Italy) every week during the study period (see Fig. 3). Before measurement, the EAA-30r-EP equipment designed to measure real-time NH₃ and the G2103 measurements from December 2020 to January 2021 were calibrated by diluting 1 ppm of standard gas. Because of the characteristics of EAA-30r-EP, it can be calibrated using only one concentration; it was calibrated by injecting 38.23 ppb, followed by sequential injections of 29.36, 19.66, 9.85, and 0 ppb after calibration to ensure the reliability of the measured values at low concentrations, as shown in Fig. 4 (slope: 1.03, intercept: -0.8). Unlike EAA-30r-EP, G2103 can be calibrated using various concentrations; G2103 was calibrated using 38.23, 29.36, 19.66, 9.85, and 0 ppb injections into the EAA-30r-EP, followed by 29.36, 9.85, and 0 ppb reinjections after calibration, to ensure reliability even at low concentrations (slope: 0.99, intercept: 0.41).

The passive sampler collected samples every week, using 10 mL of ultrapure water (DI water) for elution, after which the NH₄⁺ concentration was measured using

Dionex Aquion (Thermo Fisher Scientific, USA) equipment. Dionex Aquion utilizes ion chromatography through a 20 mM mobile phase of CH_3SO_3H , a CS12A column, and a guard column of CG12A. The concentration of atmospheric NH_3 was calculated using Eq. 1.

$$D_{\text{NH}_{5}}(T,P) = D_{0.1} \times \left(\frac{P_{0}}{P}\right) \times \left(\frac{T}{T_{0}}\right)^{1.81}$$

$$\tag{1}$$

where D_{NH_3} is the diffusivity of NH₃ in the atmosphere, calculated using the $D_{0.1} = 0.1978 \text{ cm}^2 \text{ s}^{-1}$ when the temperature (*T*) and pressure (*P*) at the time of measurement are 0°C and 1 atm, respectively (Massman *et al.*, 1998). The adsorption flow rate (Q_{NH_3}) in the passive sampler was calculated using Eq. 2.

$$Q_{NH_3} = D_{NH_3}(T, P) \times \frac{A}{\Delta X}$$
(2)

where A is the cross-sectional adsorption area available to the passive sample, and ΔX is the diffusion distance of the sampler. According to a previous study (Puchalski *et al.*, 2011), $A/\Delta X$ values of the NH₃ passive sampler of Radiello used in this study were reported to be 14.2 cm. Q_{NH_3} was used to calculate the concentration of atmospheric NH₃ ($C_{NH_3, passive}$), and the mass of NH₃ (m_{NH_3}) calculated by ion chromatography (IC) using Eq. 3 is as follows:

$$C_{NH_{s}, passive} = \frac{10^{\circ} \times m_{NH_{s}}}{t \times Q_{NH_{s}}}$$
(3)

Samples were collected every alternate week using an annual denuder and phosphorous acid; a capture material, such as a passive sampler was dissolved in methanol to coat the denuder. They were extracted using 10 mL of ultrapure liquid and analyzed using the same method as that of a passive sampler, and the concentration of NH_4^+ in the atmosphere (C_{NH_y} denuder) was calculated using Eq. 4:



Fig. 3. The schematic diagram for measuring ambient NH₃ concentration.



Fig. 4. Post-calibration curve of real-time NH_3 measuring instruments.

$$C_{NH_{sr}\,denuder} = \frac{0.945 \times c_{NH_{4}} \times V_{ext}}{V_{act.\,air}} \tag{4}$$

where 0.945 is the ratio of the weights of NH_3 and NH_4^+ , $c_{_{NH_4^+}}$ is the concentration of NH_4^+ measured by ion chromatography (in $\mu g m^{-3}$), V_{ext} is the amount of ultrapure

Ammonia Measurement in Varying Inlet Tubes



Fig. 5. Absorption ratio of NH₃ on different test tube materials.

liquid used for the denuder extraction (in L), and $V_{act. air}$ is the amount of sample collected from the atmosphere (L).

3. RESULTS

3.1 NH₃ Adsorption Depending on Moisture and Inlet Conditions

Upon comparing the degree of adsorption according to the different tube materials, the concentration of NH₃ absorbed by SUS was the highest (27.53%) when a 1 m tube was used without moisture, followed by PFA at approximately 4.08% (see Fig. 5); PVDF and PTFE did not show NH₃ adsorption. In the 2 m experimental tube, the SUS tube was excluded because it showed the highest adsorption; the adsorption onto other materials was in the descending order of PFA (approximately 6.56%), PVDF (approximately 2.67%), and PTFE (approximately 1.03%). Based on these results, the PTFE tube was found to have the lowest degree of NH₃ adsorption; thus, further experiments in this study on humidity, length, and inlet tube heating using a PTFE tube need to be conducted.

The concentration of NH₃, according to the humidity divided by the length of the PTFE tube, was stabilized at 100 ppb for a similar time. The adsorption ratio of all the tubes measured was very low, that is, 1%. The differences in concentration reduction were due to the moisture influencing the NH₃ concentration, which momentarily dropped immediately after NH₃ was injected (see Fig. 6). It was judged that regardless of the length, the degree of instantaneous adsorption of NH₃ decreased with increasing moisture. It was found that moisture reduces the adsorption of NH₃ at an appropriate humidity level 🔘 AJAE 🛛 Asian Journal of Atmospheric Environment, Vol. 15, No. 4, 2021139, 2021



Fig. 6. The maximum loss of NH₃ at different lengths and RHs of each test tube.



Fig. 7. Timeline of NH₃ adsorption by a heating test tube.

(Vaittinen *et al.*, 2018). It was confirmed that NH_3 adsorption at high moisture could be measured more accurately through real-time measurements of the general air quality. However, as the length of the tube increased, the degree of NH_3 adsorption increased. The decrease in NH_3 concentration according to the length was found to be 1.74% m⁻¹ at 80% moisture, 2.94% m⁻¹ at 60% moisture, 3.72% m⁻¹ at 40% moisture, and 7.48% m⁻¹ at 20% moisture. Thus, it was concluded that the length of the inlet tube affected the real-time NH_3 concentration measurements, especially in areas with low moisture.

Fig. 7 shows the change in the concentration measured when the heating was repeatedly turned off and on in the experimental tube to investigate whether there is a difference in NH₃ adsorption and if the inlet tube is heated or not. Neuman (1999) reported that the adsorption of HNO₃ in the sampling tube, which is similar to the chemical properties of NH₃ (sticky molecules), was not

30°C. In this test, 45°C was chosen because the box and cavity temperatures of G2103 are usually 45°C. When 100 ppb NH₃ was injected into the experimental tube, it was measured to be 90.96 ppb without heating. However, it rapidly increased to 100.31 ppb during heating. Thus, the concentration of NH₃ adsorbed to the experimental tube increased as the degree of adsorption in the tube was reduced by heating compared to that without heating. **3.2 Measurement of Atmospheric NH₃ Using**

observed when the tube temperature was higher than

3.2 Measurement of Atmospheric NH₃ Using Real-time and Non-real-time Measurement Methods

Real-time NH₃, measured using EAA-30r-EP and G2103 from May 19, 2020, to July 12, 2020, at the Honam Air and Environment Research Institute, was prepared by manufacturing a PVDF tube, conventionally announcing it to have the lowest adsorption of NH₃ (Vaittinen et al., 2018). The length of the inlet tube was 2 m, and the tube was heated to 45°C. At that time, EAA-30r-EP and G2103 were calibrated using different calibration devices and gases. Before June 5, measurements were performed with the heater of the inlet tube of EAA-30r-EP turned off to confirm the difference in ammonia concentration measured when the inlet tube was not heated and when it was heated. For EAA-30r-EP and G2103, NH₃ measurements were performed between December 5, 2020, and January 3, 2021, and these were manufactured using PTFE inlet tubes, that is, those measured to have the lowest degree of adsorption in section "3.1 NH₃ adsorptions depending on moisture and inlet



Fig. 8. Timeline of NH_3 measurement of ambient in Gwang-ju from (a) May 2020 to July 2020. The EAA-30r-EP inlet was not heated from May 5 to June 5, and the two instruments were calibrated separately. (b) Measurement between December 2020 and January 2021 after completing calibration by the same system.

tube conditions" Furthermore, measurements were performed after simultaneous calibration using the same calibration device and calibration gas.

From May 19, 2020, to July 12, 2020, the average concentration of NH₃ measured using the EAA-30r-EP realtime measuring equipment was $10.09 \text{ ppb} (\pm 6.74 \text{ ppb})$ and that by using the G2103 equipment was 16.11 ppb $(\pm 6.82 \text{ ppb})$. The passive sampler measured data every week during the same period. The average value was 10.74 ppb (\pm 3.84 ppb); the annual denuder averaged 11.41 ppb (\pm 4.06 ppb). The average concentration of NH₃, measured from December 5, 2020, to January 3, 2021, was 9.54 ppb (\pm 6.98 ppb) for the EAA-30r-EP, 9.46 ppb (\pm 7.16 ppb) for the G2103, 8.4 ppb (\pm 2.95 ppb) for the passive sampler, and $5.74 \text{ ppb} (\pm 0.29 \text{ ppb})$ for annual denuder (see Fig. 8). The standard deviation of the passive sampler and annual denuder was lower than that of real-time analysis equipment because the concentrations calculated from the sample collected for one week cannot be reflected in the change in NH₃ concentration, which changes in real-time within a certain



Fig. 9. Comparison of various types of NH_3 measurement. (a) Comparison of real-time NH_3 measurement by heating or not heating the inlet (May 2020–July 2020). (b) Comparison of two types of real-time NH_3 measuring instruments (December 2020–January 2021). (c) Comparison of annual denuder with a passive sampler (May 2020–July 2020 and December 2020–January 2021), and (d) Comparison of real-time measurements with the passive sampler (June 2020–July 2020 and December 2020–January 2021).

range. For the annual denuder, samples were collected every alternate week. Hence, the concentration was lower than that measured by the passive sampler every week and other real-time measurement equipment.

Fig. 9 shows the correlation between the analysis methods both in real-time and non-real-time during the same period. To minimize the change in concentration due to temperature differences over time, the results of real-time NH₃ concentration were averaged to 1 h data and then converted from ppb into μ g m⁻³ using atmospheric temperature and pressure measurements by the Honam Air and Environment Research Institute. Fig. 9 a is based on comparisons between the two types of real-time measuring equipment used between May and July 2020; the EAA-30r-EP, without heating the inlet tube, was approximately 39% lower than G2103 with a heated inlet tube. When EAA-30r-EP is heated, it measures approximately 29% lower than G2103; a 10% reduction

in the difference was observed based on whether the inlet tube was heated or not, indicating that heating the inlet tube during the real-time measurement is necessary. Moreover, based on comparisons of NH₃ concentrations measured in real-time using different calibration devices and gases, the coefficient of determination was found to be 0.94, indicating dispersed correlation when inlet tubes were used under similar conditions. To investigate whether the dispersed correlation resulted from using other equipment for measurements between December 2020 and January 2021, the aforementioned real-time measurement devices were calibrated using the same calibration device and gas (Fig. 9, b). EAA-30r-EP was found to be 0.5% lower than that of G2103, and the coefficient of determination was 0.99, that is, a very high correlation. It has been confirmed that the calibration method and the heating of the inlet tube are very important for real-time NH₃ measurements. Fig. 9, c shows that the annual denuder collecting data biweekly for one week was 2% higher than the passive sampler collecting NH₃ during the same period. The coefficient of determination was calculated to be 0.996, showing a very high correlation between the two non-real-time NH₃ measurements. Based on the comparisons and analysis of the NH₃ concentration measured by the passive sampler collected every week during the measurement period and real-time equipment with a heated inlet tube (June 5, 2020, to July 12, 2020, and December 5, 2020, to January 3, 2021) (see Fig. 9, d), the concentration of NH_3 measured using real-time equipment and non-real-time measurement methods was correlated at 0.98, with a difference of 7%.

4. CONCLUSIONS

This study evaluated NH_3 adsorption in inlet tubes used in real-time NH_3 measurements according to the material, length, humidity, and heating of the inlet tube. Furthermore, atmospheric NH_3 was measured using the inlet tube with the lowest adsorption. It was compared and analyzed with the concentration measured in the passive sampler and annual denuder, non-real-time measurement methods. The materials used to make tubes for the inlet tube experiment were SUS316L, PFA, PTFE, and PVDF; the injection concentration of NH_3 was stabilized at 100 ppb. The absorption rate of NH_3 in the PTFE tube was measured to be the lowest (i.e., 1.03%) at 2 m when using the PTFE tube. Based on these results, the NH_3 adsorption, according to length and moisture, measured using the PTFE tube was the lowest, the relative humidity increased from 20% to 80%, and the degree of instantaneous adsorption decreased from 7.48 ppb m⁻¹ to 1.74 ppb m⁻¹; indicating that the moisture reduces NH_3 adsorption to the inlet tube. Furthermore, the degree of adsorption to the tube was reduced when the experimental tube was heated. It was proven to be more effective at low humidity.

From May 2020 to July 2020 and December 2021 to January 2021, atmospheric NH₃ was measured using real-time techniques and non-real-time measurement methods at the Honam Air and Environment Research Institute. The results indicated that the average concentrations of NH₃ measured with real-time equipment EAA-30r-EP were 10.09 ppb and 9.54 ppb for both study periods, respectively. The concentrations measured with G2103 were 16.11 ppb and 9.46 ppb, respectively. The average concentration from the passive sampler, a nonreal-time measuring instrument, was 10.74 ppb and 8.4 ppb, respectively. When comparing the two methods, the real-time methods, EAA-30r-EP, and G2103 showed a concentration difference of 0.5%. The non-real-time measurement method using the annual denuder passive sampler, showed a difference of 2%.

Moreover, the NH₃ concentration was approximately 10% lower when the inlet tube of the real-time equipment was not heated. It was also confirmed that there is a difference in concentration according to the calibration device and gas. Overall, real-time NH₃ measurement can minimize the lack of NH₃ measurement using an inlet tube made of PTFE, a material that can be heated. It is important to provide uniformity to the calibration method, and the calibration gas used when measuring the NH₃ concentration in real-time at various places or times.

ACKNOWLEDGMENT

This study was funded by the National Institute of Environmental Research (NIER-2020-04-02-041), and additional data processing was supported by the Hankuk University of Foreign Studies Research Fund (202111 72001) and the Ministry of Education of the Republic of Korea and the National Research Foundation of Korea (NRF-2019S1A6A3A02058027).

DISCLAIMER

All the results in this study were obtained by setting the parameters for the passive sampler and denuder, according to the referenced literature.

REFERENCES

- Ellis, R., Murphy, J., Pattey, E., Haarlem, R.V., O'Brien, J., Herndon, S. (2010) Characterizing a quantum cascade tunable infrared laser differential absorption spectrometer (QC-TIL-DAS) for measurements of atmospheric ammonia. Atmospheric Measurement Techniques, 3(2), 397-406. https:// doi.org/10.5194/amt-3-397-2010
- Hsieh, L.-T., Chen, T.-C. (2010) Characteristics of ambient ammonia levels measured in three different industrial parks in southern Taiwan. Aerosol and Air Quality Research, 10(6), 596–608. https://doi.org/10.4209/aaqr.2010.06. 0044
- Jordan, C.E., Crawford, J.H., Beyersdorf, A.J., Eck, T.F., Halliday, H.S., Nault, B.A., Chang, L.-S., Park, J., Park, R., Lee, G. (2020) Investigation of factors controlling PM_{2.5} variability across the South Korean Peninsula during KORUS-AQ. Elementa: Science of the Anthropocene, 8. https:// doi.org/10.1525/elementa.424
- Junsu, P., Jaeyoun, R., Joonbum, J., Song, M. (2020) Origins and Distributions of Atmospheric Ammonia in Jeonju during 2019~2020. Journal of Korean Society for Atmospheric Environment, 36(2), 262–274, (in Korean with English abstract). https://doi.org/10.5572/KOSAE.2020.36.2.262.
- Kang, S., Park, G., Park, T., Ban, J., Kim, K., Seo, Y., Choi, J., Seo, S., Choi, J., Bae, M.-S. (2020) Semi-continuous Measurements of Water-soluble Organic Carbon and Ionic Composition of PM_{2.5} in Baengnyeong Island during the 2016 KORUS-AQ (Korea-United States Air Quality Study). Asian Journal of Atmospheric Environment, 14(3). https://doi. org/10.5572/ajae.2020.14.3.307
- Kim, N., Yum, S.S., Park, M., Park, J.S., Shin, H.J., Ahn, J.Y. (2020) Hygroscopicity of urban aerosols and its link to sizeresolved chemical composition during spring and summer in Seoul, Korea. Atmospheric Chemistry and Physics, 20(19), 11245-11262. https://doi.org/10.5194/acp-20-11245-2020
- Lee, J. (2019) Performance test of microaeth[®] AE51 at concentrations lower than 2 μg/m³ in indoor laboratory. Applied Sciences, 9(13), 2766. https://doi.org/10.3390/app9132 766
- Li, Y., Thompson, T.M., Damme, M.V., Chen, X., Benedict, K.B., Shao, Y., Day, D., Boris, A., Sullivan, A.P., Ham, J. (2017) Temporal and spatial variability of ammonia in urban and agricultural regions of northern Colorado, United States. Atmospheric Chemistry and Physics, 17(10), 6197– 6213. https://doi.org/10.5194/acp-17-6197-2017
- Massman, W. (1998) A review of the molecular diffusivities of H₂O, CO₂, CH₄, CO, O₃, SO₂, NH₃, N₂O, NO, and NO₂ in

air, O₂ and N₂ near STP. Atmospheric Environment, 32(6), 1111–1127. https://doi.org/10.1016/S1352-2310(97)00 391-9

- Neuman, J.A., Huey, L.G., Ryerson, T.B., Fahey, D.W. (1999) Study of Inlet Materials for Sampling Atmospheric Nitric Acid. Environmental Science & Technology, 33, 1133– 1136. https://doi.org/10.1021/es980767f
- Pan, Y., Tian, S., Zhao, Y., Zhang, L., Zhu, X., Gao, J., Huang, W., Zhou, Y., Song, Y., Zhang, Q. (2018) Identifying ammonia hotspots in China using a national observation network. Environmental Science & Technology, 52(7), 3926–3934. https://doi.org/10.1021/acs.est.7b05235
- Park, G., Kim, K., Kang, S., Park, T., Ban, J., Lee, T. (2019) The Chemical Characteristics and Formation of Potential Secondary Aerosol (PSA) using an Oxidation Flow Reactor (OFR) in the Summer: Focus on the Residential Area, Suwon. Journal of Korean Society for Atmospheric Environment, 35(6), 786–801, (in Korean with English abstract). https://doi.org/10.5572/KOSAE.2019.35.6.786.
- Park, T., Ban, J., Kang, S., Ghim, Y.S., Shin, H.-J., Park, J.S., Park, S.M., Moon, K.J., Lim, Y.-J., Lee, M.-D., Lee, S.-B., Kim, J., Kim, S.T., Bae, C.H., Lee, Y., Lee, T. (2018) Chemical Characteristics of PM1 using Aerosol Mass Spectrometer at Baengnyeong Island and Seoul Metropolitan Area. Journal of Korean Society for Atmospheric Environment, 34(3), 430–446, (in Korean with English abstract). https://doi. org/10.5572/KOSAE.2018.34.3.430.
- Paulot, F., Jacob, D.J., Pinder, R., Bash, J., Travis, K., Henze, D. (2014) Ammonia emissions in the United States, European Union, and China derived by high-resolution inversion of ammonium wet deposition data: Interpretation with a new agricultural emissions inventory (MASAGE_NH₃). Journal of Geophysical Research: Atmospheres, 119(7), 4343-4364. https://doi.org/10.1002/2013JD021130
- Pollack, I.B., Lindaas, J., Roscioli, J.R., Agnese, M., Permar, W., Hu, L., Fischer, E.V. (2019) Evaluation of ambient ammonia measurements from a research aircraft using a closedpath QC-TILDAS operated with active continuous passivation. Atmospheric Measurement Techniques, 12(7), 3717-3742. https://doi.org/10.5194/amt-12-3717-2019
- Sharma, S.K., Datta, A., Saud, T., Saxena, M., Mandal, T., Ahammed, Y., Arya, B. (2010) Seasonal variability of ambient NH₃, NO, NO₂ and SO₂ over Delhi. Journal of Environmental Sciences, 22(7), 1023–1028. https://doi.org/10. 1016/S1001-0742(09)60213-8
- Bobrutzki, K., Braban, C., Famulari, D., Jones, S., Blackall, T., Smith, T., Blom, M., Coe, H., Gallagher, M., Ghalaieny, M. (2010) Field inter-comparison of eleven atmospheric ammonia measurement techniques. Atmospheric Measurement Techniques, 3, 91–112. https://doi.org/10.5194/amt-3-91-2010
- Yi, X., Zhang, Z., Smith, P. (2021) Real-time measurements of landfill atmospheric ammonia using mobile white cell differential optical absorption spectroscopy system and engineering applications. Journal of the Air & Waste Management Association, 71(1), 34-45. https://doi.org/10.1080 /10962247.2020.1820405
- Zhou, C., Zhou, H., Holsen, T.M., Hopke, P.K., Edgerton, E.S.,

OAJAE Asian Journal of Atmospheric Environment, Vol. 15, No. 4, 2021139, 2021

Schwab, J.J. (2019) Ambient ammonia concentrations across New York state. Journal of Geophysical Research: Atmospheres, 124(14), 8287–8302. https://doi.org/10.1029/ 2019JD030380

Zhu, L., Henze, D.K., Bash, J.O., Cady-Pereira, K.E., Shephard,

M.W., Luo, M., Capps, S.L. (2015) Sources and impacts of atmospheric NH 3: current understanding and frontiers for modeling, measurements, and remote sensing in North America. Current Pollution Reports, 1(2), 95–116. https://doi.org/10.1007/s40726-015-0010-4